

project: Wind monitoring
 instrument: EKO21N /iBOX
 ser. Nr/code: 152
 date: April 2009
 range: 0 to 50 m/s
 channel: 2
 FIRST CLASS ANEMOMETER CALIBRATION
 calibration data anemometer
 calibration accuracy: 0.1 m/s
 anemometer serial nr: 0856_09
 (refer calibration certificate)
 a= 0.76580 (slope)
 b= 0.271 (m/s) (offset)

windspeed v (m/s)= a * frequency (Hz) + b (if $v > 0$)

Calibration of EKO 21:

anemometer frequency (Hz)	wind speed calibration certificate (m/s)	indicated windspeed EKO 21 (m/s)	deviation (m/s)
5.00	4.10	4.13	-0.03
10.00	7.93	7.95	-0.02
15.00	11.76	11.78	-0.02
20.00	15.59	15.62	-0.03
25.00	19.42	19.45	-0.03

Authorization by EKOPOWER:

S Brans (calibration Officer)